



WASHINGTON STATE PATROL
SCALE TEST REPORT AND CERTIFICATION OF INSPECTION

DATE: Mar 15, 2006 INITIALS: DEO OWNER: W/SR
DISTRICT: 87 ADDRESS: _____
LOCATION: 7th Ship COUNTY: _____
MAKE: Huenn TYPE: port wheel weigher ACCURACY CLASS: III
MODEL: 101 ID. NO.: 23110 CAPACITY: 20,000 lbs SECTIONS: 1
MINIMUM GRADUATIONS: 50 lb REG. ELEMENT: dial PLATFORM: 15" X 26"


| LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR | LOAD POSITION | TEST WEIGHT | SCALE INDICATION | ERROR |
|---------------|--------------|------------------|------------|---------------|-------------|------------------|-------|
| <u>Center</u> | <u>5000</u> | <u>5000</u> | <u>0</u> | | | | |
| <u>"</u> | <u>10000</u> | <u>10000</u> | <u>0</u> | | | | |
| <u>"</u> | <u>15000</u> | <u>15050</u> | <u>+50</u> | | | | |
| <u>Center</u> | <u>20000</u> | <u>20000</u> | <u>0</u> | | | | |
| | | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | | | | | | | |
| | | | | | | | |

Using testing procedures set forth in Handbook 44 promulgated by the National Institute of Standards and Technology and test weights which were certified as accurate as is shown on the attached "Certificate of Accuracy", No. 19577

I do hereby certify under penalty of perjury under the laws of the state of Washington that the above described weighing device met or exceeded the standards of accuracy.

Signed by: David Cronin Date: 3/28/06

State of Washington
County of Thurston
Signed and sworn to (or affirmed) before
me on March 28, 2006 by
Michael J. Evans
(Signature of notary public)
Michael J. EVANS
Printed Name of Notary
My appointment expires 12-27-07



3000-362-021 (R) (10/98)

DISTRIBUTION: WHITE - COURT GREEN - DISTRICT CANARY - PROPERTY MANAGEMENT PINK & GOLDENROD - SERVICE TRUCKS